


**Lampiran 1: Prosedur Ekstraksi Isoflavon dari Tauco**

## Lampiran 2: Analisis kuantitatif Isoflavon dengan KCKT

### Perhitungan konsentrasi daidzein dan genestein dalam 100 g sampel

$A_x$  = area rata-rata yang didapatkan dari puncak isoflavon

$A_s$  = area rata-rata dari standar isoflavon

$m_s$  = massa isoflavon standar ( $\mu\text{g}$ )

$m_x$  = massa isoflavon ( $\mu\text{g}$ )

$C$  = konstanta volume terhadap jalur kerja

Perhitungan yang dipakai adalah:

$$m_x = C \times \frac{A_x}{A_s} \times m_s$$

Konsentrasi isoflavon adalah  $m_x/100$  g sampel

Perubahan volume sampel terhadap jalur kerja:

$$a \text{ mL} \rightarrow b \text{ mL} \Rightarrow c \text{ mL}$$

Keterangan:

Tanda ( $\rightarrow$ ) artinya diencerkan sampai b mL dari a mL

Tanda ( $\Rightarrow$ ) artinya diambil c mL dari b mL

#### A. Konsentrasi daidzein dan genestein dalam 100 g tempe

Perubahan volume sampel terhadap jalur kerja:

$$1 \text{ mL} \rightarrow 10 \text{ mL} \Rightarrow 1 \text{ mL} \rightarrow 5 \text{ mL} \Rightarrow 2 \text{ mL} \rightarrow 4 \text{ mL} \Rightarrow 2,3 \text{ mL} \rightarrow 40 \text{ mL} \Rightarrow 20 \mu\text{L}$$

##### a. Konsentrasi daidzein

$$A_{\text{daidzein}} = 63750 \text{ satuan luas}$$

$$A_{s, \text{ daidzein}} = 651876,5 \text{ satuan luas}$$

$$m_{s, \text{ daidzein}} = 0,2 \mu\text{g}$$

$$\begin{aligned}
 m_{\text{daidzein}} &= [(10/1) \times (10/1) \times (5/1) \times (5/2) \times (4/2) \times (4/2,3) \times (40/2,3) \times \\
 &\quad (40000/20)] \times (63750/651876,5) \times 0,2 \mu\text{g} \\
 m_{\text{daidzein}} &= 2957870,548 \mu\text{g} \\
 m_{\text{daidzein}} &= 2957,8705 \text{ mg}
 \end{aligned}$$

Konsentrasi daidzein adalah 2957,8705 mg/100 g tempe

#### b. Konsentrasi genestein

$$\begin{aligned}
 A_{\text{genestein}} &= 33284,5 \text{ satuan luas} \\
 A_{\text{s, genestein}} &= 2069868 \text{ satuan luas} \\
 m_{\text{s, genestein}} &= 0,2 \mu\text{g} \\
 m_{\text{genestein}} &= [(10/1) \times (10/1) \times (5/1) \times (5/2) \times (4/2) \times (4/2,3) \times (40/2,3) \times \\
 &\quad (40000/20)] \times (33284,5/2069868) \times 0,2 \mu\text{g} \\
 m_{\text{genestein}} &= 486366,5 \mu\text{g} \\
 m_{\text{genestein}} &= 486,3665 \text{ mg}
 \end{aligned}$$

Konsentrasi genestein adalah 486,3665 mg/100 g tempe

#### B. Konsentrasi daidzein dan genestein dalam 100 g tauco

Perubahan konsentrasi sampel terhadap jalur kerja:

$$1\text{ mL} \rightarrow 12\text{ mL} \Rightarrow 2\text{ mL} \rightarrow 4\text{ mL} \Rightarrow 3,3\text{ mL} \rightarrow 60\text{ mL} \Rightarrow 20\text{ }\mu\text{L}$$

##### a. Konsentrasi daidzein

$$\begin{aligned}
 A_{\text{daidzein}} &= 103660,5 \text{ satuan luas} \\
 A_{\text{s, daidzein}} &= 651876,5 \text{ satuan luas} \\
 m_{\text{s, daidzein}} &= 0,2 \mu\text{g} \\
 m_{\text{daidzein}} &= [(12/1) \times (12/2) \times (4/2) \times (4/3,3) \times (60/3,3) \times 60000/20] \times \\
 &\quad (103660,5/651876,5) \times 0,2 \mu\text{g}
 \end{aligned}$$

$$m_{\text{daidzein}} = 302792,5 \mu\text{g}$$

$$m_{\text{daidzein}} = 302,7925 \text{ mg}$$

Konsentrasi daidzein adalah 302,7925 mg/100 g tauco

**b. Konsentrasi genestein**

$$A_{\text{genestein}} = 52620 \text{ satuan luas}$$

$$A_{\text{s, genestein}} = 2069868 \text{ satuan luas}$$


$$m_{\text{s, genestein}} = 0,2 \mu\text{g}$$


$$m_{\text{genestein}} = [(12/1) \times (12/2) \times (4/2) \times (4/3,3) \times (60/3,3) \times (60000/20)] \times (52620/2069868) \times 0,2 \mu\text{g}$$

$$m_{\text{genestein}} = 48406,7 \mu\text{g}$$

$$m_{\text{genestein}} = 48,4067 \text{ mg}$$


Konsentrasi genestein adalah 48,4067 mg/100 g tauco



Lampiran 3: Isoflavon dalam Kedelai<sup>[6]</sup>

(a)

No.	Senyawa Isoflavon	R <sub>1</sub>	R <sub>2</sub>	R <sub>3</sub>
1	Daidzin	H	H	H
2	Glisitin	H	OCH <sub>3</sub>	H
3	Genistin	H	H	OH
4	6''-O-Malonil daidzin	COCH <sub>2</sub> COOH	H	H
5	6''-O-Malonil glisitin	COCH <sub>2</sub> COOH	OCH <sub>3</sub>	H
6	6''-O-Malonil genistin	COCH <sub>2</sub> COOH	H	OH
7	6''-O-Asetil daidzin	COCH <sub>3</sub>	H	H
8	6''-O-Asetil glisitin	COCH <sub>3</sub>	OCH <sub>3</sub>	H
9	6''-O-Asetil genistin	COCH <sub>3</sub>	H	OH



(b)

Daidzein: R<sub>1</sub> = OH; R<sub>2</sub> = H; R<sub>3</sub> = H; R<sub>4</sub> = OH

Genestein: R<sub>1</sub> = OH; R<sub>2</sub> = H; R<sub>3</sub> = OH; R<sub>4</sub> = OH

(a) Struktur isoflavon glikosida; (b) Struktur isoflavon aglikon

### D-7000 HPLC System Manager Report

Analyzed: 10/31/00 10:46 AM

Reported: 10/31/00 11:46 AM

Processed: 10/31/00 11:46 AM

Data Path: C:\WINS2APP\HSM\samples\DATA\0022\

Processing Method: tmh

System(acquisition): Sys 1

Series:0022

Application: SAMPLES

Vial Number: 1

Sample Name: tmh001


Vial Type: UNR

Injection from this vial: 1 of 1

Volume: 20.0 ul

sample description:

Chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

column type:

developed by: ratna

Pump A Type: L-7100

Solvent A: air/acet

Pump B Type: L-7100

Solvent B: meoh/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc i	BC
1	2.80	115060	1.359	BB
2	2.93	28660	0.331	BB
3	3.36	520796	6.152	BV
4	8.86	115869	1.369	VV
5	10.00	79563	0.940	VV
6	11.09	112768	1.332	VV
7	14.64	164718	1.946	VV
8	15.65	61595	0.728	VV
9	16.85	42870	0.506	VV
10	18.32	73577	0.893	VB
11	26.77	12903	0.152	BB
12	28.61	19236	0.227	BB
13	31.97	30671	0.362	BB
14	34.00 ✓	649008	7.667	BB
15	36.96 ✓	2059721	24.533	BV

D 2 G

D-7000 HSM: SAMPLES		Series: 0022	Report: original	System: Sys 1
16	41.71	211346	2.497	VV
17	44.37	331175	6.275	VV
18	46.53	159270	1.662	VV
19	47.60	123414	1.438	VV
20	48.67	122047	1.442	VV
21	50.40	252065	2.978	VV
22	51.47	205666	2.432	VV
23	52.51	115667	1.366	VV
24	53.41	161934	1.913	VV
25	54.40	223876	2.645	VV
26	55.52	310043	3.663	VV
27	55.64	122301	1.445	VV
28	56.59	334251	3.949	VV
29	56.83	405877	4.793	VV
30	57.44	505859	5.976	VV
31	58.32	452731	5.946	VV
32	59.41	136653	1.636	VV
		8464810	100,000	

peak rejection level: 0



st 116

D-7000 HSM: SAMPLES

Series: 0023

Report: original

System: Sys 1

### D-7000 HPLC System Manager Report

Analyzed: 10/31/00 12:16 PM


Reported: 10/31/00 01:28 PM  
 Processed: 10/31/00 01:16 PM

Data Path: C:\WIN32APP\HSM\samples\DATA\0023\  
 Processing Method: tmh

system(acquisition): Sys 1  
 Application: SAMPLES  
 Sample Name: tmh002  
 Injection from this vial: 1 of 1  
 sample description:

Series: 0023  
 Vial Number: 2  
 Vial Type: UNK  
 volume: 20.0 ul

Chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh  
 Column type:  
 Pump A Type: L-7100  
 Solvent A: air/acet  
 Pump B Type: L-7100  
 Solvent A: meoh/acet  
 Peak Quantitation: AREA  
 Calculation Method: AREA\*

developed by: ratna

NO.	RT	Area	Conc 1	BC
1	2.83	144380	1.709	BB
2	2.96	31502	0.373	BB
3	3.39	253217	2.997	BB
4	8.99	22400	0.265	BB
5	10.11	15640	0.185	BB
6	11.83	23538	0.279	BB
7	13.41	33263	0.354	BB
8	16.69	14109	0.167	BB
9	17.15	11469	0.136	BB
10	20.40	29916	0.354	BB
11	28.80	11500	0.136	BB
12	30.59	11380	0.135	BB
13	32.16	31220	0.369	BB
14	34.32	654745	7.749	BB
15	37.26	2060015	24.616	BB



3

D-7000 HSM: SAMPLES		Series: 0023	Report: original	System: Sys 1	
16	44.83	800107		9.470	BV
17	46.61	131327		1.557	VV
18	47.76	179926		2.129	VV
19	48.72	139789		1.654	VV
20	50.56	296364		3.507	VV
21	51.49	254672		3.016	VV
22	52.75	162654		2.162	VV
23	53.60	151663		1.795	VV
24	54.61	266383		3.153	VV
25	55.92	534158		6.322	VV
26	56.72	394145		4.665	VV
27	56.99	466664		5.526	VV
28	57.57	666916		7.693	VV
29	58.43	615643		7.288	VB
		8449269		100.000	

Peak rejection level: 0



7011

D-7000 HSM: SAMPLES

Series: 0025

Report: original

System: Sys 1

### D-7000 HPLC System Manager Report

Analyzed: 10/31/00 02:40 PM

Reported: 10/31/00 03:26 PM

Processed: 10/31/00 03:26 PM

Data Path: C:\WIN32APP\HSM\samples\DATA\0025\

Processing Method: tmh

System(acquisition): Sys 1

series:0025

Application: SAMPLES

Vial Number: 1

Sample Name: tmh001


Vial Type: UNN

Injection from this vial: 1 of 1

volume: 20.0 ul

sample description:

chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

column type:

Developed by: ratna

Pump A Type: L-7100

Solvent A: air/acet

Pump B Type: L-7100

Solvent B: meoh/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc 1	BC
1	2.05	28867	2.260	BV
2	2.77	74592	5.872	VV
3	3.01	3662	0.287	TBV
4	3.20	9677	0.758	IVB
5	3.47	76187	5.965	BV
6	5.41	200503	15.698	VV
7	8.80	190218	14.853	VB
8	11.23	14845	1.163	BB
9	13.49	10640	0.833	BB
10	16.96	3460	0.271	BB
11	19.17	58783	4.601	BB
12	22.43	34311	2.686	BB
13	24.35	24725	1.936	BB
14	29.52	17660	1.383	BB
15	32.32	52200	2.521	BB

D-7000 HSM: SAMPLES		Series: 0025	Report: original	System: Sys 1
16	35.44 v	103970	8.140	BB
17	36.60	30569	2.393	BB
18	36.24 v	52520	4.112	BB
19	39.95	8800	0.689	BB
20	41.41	60776	4.758	BV
21	45.52	226994	17.773	VV
22	47.41	12869	1.008	VB
			1277212	100,000

Peak rejection level: 0



D-7000 HSM: SAMPLES

Series: 0023

Report: original

System: Sys 1

channel 1 Noise: Not Measured

channel 1 drift: Not measured

Configuration parameters:

interface module: D-7000

channel 1 detector: L-7450

Column Oven: None

Pump A: L-7100

Pump C: None

column name:

gradient mode: high

channel 2 detector: None

Autosampler: Manual

Pump B: L-7100

External Instrument Software: None

Method information:

Method Name: tmh

Description:

analisaas

Developed by: ratna

Pump Setup:

All Pumps Pressure Limit: 0 to 420 kgf/cm2

Pump A (L-7100):

solvent A: air/acet

Pump B (L-7100):

Solvent B: meoh/acet

pump table

Time (min)	%air/a	%meoh/	Total Flow (ml/min)	Event 1	Event 2	Event 3	Event 4
0.0	100.0	0.0	1.000				
5.0	100.0	0.0	1.000				
55.0	0.0	100.0	1.000				
60.0	100.0	0.0	1.000				

Channel 1 Detector (L-7450):

Spectral Bandwidth: 4 nm

Absorbance Mode: Normal (2.0AU)

wavelength range: 190 to 890 nm

start time: 0.00 min

Spectral Interval: 1000 msec

Auto Zero before Injection: YES

monitoring wavelength: 260 nm

stop time: 60.00 min

Method DF for channel 1

calculation method:

calculation method: AREAS

Peak identification window: % Time

UNK peaks identification rule: Closest peak

Update RT in component Table: NO

Do library search: NO

Peak quantitation: Area

concentration data from method.

Do blank subtraction: NO

Component Table

RT (min)	Window (%)	Func1	Func2	Func3	E-Conc	Tolerance (%)
5.00	10.00					

Integration table

time (min)	function	value/status
0.00	NOISE	10
0.00	SMOOTHING	OFF
0.00	SENSITIVITY	20
0.00	N-METHOD	0

## DAD Processing Setup:

Peak spectrum integration enabled: NO  
Chromatogram to create: fixed at 280 nm

Peak purity check enabled: NO  
Delete DAD data after reporting: NO

## DAD Display Format:

Time range: 0.00 to 60.00 min  
Offset: 0.0  
Auto Mark Peak NL: NO  
3-D resolution: Medium  
3-D rotation: 30  
Display spectra only: NO

Absorbance scale (Au): Auto  
Wavelength range: 190 to 390 nm  
Spectrum display: Absorbance  
Auto BG Subtraction: NO  
3-D tilt: 30  
3-D mirror: NO  
Report spectra: peak top and sides.

Perform system suitability test : NO  
Perform module performance test : NO  
Perform data diagnosis : NO

## Chromatogram Display Format:

Autoscale Time Range: 0.00 to 60.00 min  
Use alternate scale: NO  
Scale to Full Chrom Time Range: YES  
Baseline overlay: YES  
Marker-in signals: NO  
Show integration time table: NO  
Show gradient curves: NO  
Report channel 1 labels in the chromatogram overlay graph.  
Picture in picture: none

Autoscale: YES  
Auto zero: NO  
Peak rejection level: 0 uv + sec  
Peak start-end markers: YES  
Peak labels: time, none  
Show sampling period time table: NO

## Report Format:

Use primary layout: YES  
Print primary layout report: YES  
Acquisition DDE: NO  
Reprocess DDE: NO  
Concentration 1 Unit: Other  
Concentration 1 scale factor: 1.0000  
Concentration 1 divide by sample amount: NO  
Concentration 2 Unit: other  
Concentration 2 scale factor: 1.0000  
Concentration 2 use component multiplier: NO  
Injection report column 1 header: PK-NUM  
Injection report column 2 header: RT  
Injection report column 3 header: AREA  
Injection report column 4 header: CONC1  
Injection report column 5 header: BC

Reported peaks: All Peaks  
Use secondary layout: NO  
Print secondary layout report: NO  
Acquisition macro name:  
Reprocess macro name:  
Concentration 1 name:  
Concentration 2 name:

### D-7000 HPLC System Manager Report

Analyzed: 11/01/00 11:01 AM

Reported: 11/01/00 12:02 PM

Processed: 11/01/00 12:02 PM

Data Path: C:\WIN32APP\HSM\samples\DATA\0028\

Processing Method: tmh

System(acquisition): Sys 1

Series:0028

Application: SAMPLES

Vial Number: 1

Sample Name: tmh001


Vial Type: UNR

Injection from this vial: 1 of 1

Volume: 20.0 ul

Sample Description:

chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

column type:

developed by: ratna

Pump A Type: L-7150

Solvent A: air/acet

pump B type: L-7100

Solvent A: meoh/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc 1	BC
1	2.77	93360	1.776	BV
2	3.04	2729	0.052	TBB
3	3.23	8790	0.167	TBB
4	3.49	247454	4.706	BV
5	5.39	22060	0.420	TBB
6	8.77	142905	2.719	VB
7	13.84	16280	0.310	BB
8	17.23	5540	0.105	BB
9	18.27	9636	0.183	BB
10	19.49	56243	0.690	BB
11	22.91	22560	0.429	BB
12	24.80	18780	0.357	BB
13	28.40	32160	0.612	BB
14	30.16	8443	0.161	BB
15	32.69	52556	1.000	BB

D-7000 HSM: SAMPLES		Series: 0028	Report: original	System: Sys 1	
16	35.79 ✓	103351	1.967		BB
17	37.15	26648	0.511		BB
18	38.53 ✓	52720	1.003		BB
19	43.79	121072	2.304		BV
20	45.92	361930	6.867		VV
21	47.65	92284	1.756		VV
22	48.72	90562	1.724		VV
23	52.21	332176	6.320		VV
24	53.84	243765	4.638		VV
25	56.53	505592	17.231		VV
26	57.17	497144	9.459		VV
27	57.44	456139	6.337		VV
28	58.00	1045656	19.896		VV
29	59.49	224812	4.278		VB
		5255389	100.000		

Peak rejection level: 0



Channel 1 Noise: Not Measured

Channel 1 drift: Not Measured

## Configuration parameters:

Interface Module: D-7000

Channel 1 Detector: L-7450

Column Oven: None

Pump A: L-7100

Pump C: None

Column Name:

Gradient mode: High

Channel 2 Detector: None

Autosampler: Manual

Pump B: L-7100

External Instrument Software: None

## Method Information:

Method Name: tmt

Description:

analisaas

Developed by: ratna

## Pump Setup:

All Pumps Pressure Limit: 0 to 420 kgf/cm<sup>2</sup>

## Pump A (L-7100):

Solvent A: air/acet

## Pump B (L-7100):

Solvent B: mech/acet

## Pump Table

Time (min)	%air/a	%mech/	Total Flow (ml/min)	Event 1	Event 2	Event 3	Event 4
0.0	100.0	0.0	1.000				
5.0	100.0	0.0	1.000				
55.0	0.0	100.0	1.000				
60.0	100.0	0.0	1.000				

## Channel 1 Detector (L-7450):

Spectral Bandwidth: 4 nm

Absorbance Mode: Normal (2.0AU)

Wavelength range: 190 to 390 nm

Start time: 0.00 min

Spectral Interval: 1800 msec

Auto Zero before Injection: YES

Monitoring wavelength: 260 nm

Stop time: 60.00 min

## Method DF for channel 1

## Calculation method:

Calculation Method: AREAS

Peak Identification Window: % Time

UNK peaks identification rule: Closest peak

Update RT in component table: NO

Do library search: NO

Peak quantitation: Area

Concentration data from method.

## Component Table

RT (min)	Window (%)	Func1	Func2	Func3	S-Conc	Tolerance (%)
5.00	10.00					

## Integration table



SPL 7c H

D-7000 HSM: SAMPLES

Series: 0029

Report: original

System: Sys 1

### D-7000 HPLC System Manager Report

Analyzed: 11/01/00 12:12 PM

Reported: 11/01/00 03:52 PM

Processed: 11/01/00 01:13 PM

Data Path: C:\WIN32APP\HSM\samples\DATA\0029\

Processing method: tmh

System(acquisition): Sys 1

Series:0029

Application: SAMPLES

Vial Number: 2

Sample Name: tmh002


Vial Type: UNR

Injection from this vial: 1 of 1

Volume: 20.0 ul

Sample description:

chrom type: mixed WL chromatogram, 260 nm



Acquisition Method: tmh

Column type:

developed by: ratna

Pump A Type: L-7100

Solvent A: air/acet

Pump B type: L-7100

Solvent A: meoh/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc 1	BC
1	2.91	178340	2.383	BB
2	3.01	22063	0.295	BV
3	3.31	14157	0.189	VB
4	3.52	315272	4.213	BV
5	5.57	8440	0.113	TBB
6	9.12	194437	2.598	VB
7	10.59	7785	0.104	BB
8	17.28	13340	0.178	BB
9	18.27	10840	0.145	BB
10	19.95	27919	0.365	BB
11	23.97	20971	0.280	BB
12	25.09	13466	0.180	BB
13	27.07	8559	0.114	BB
14	28.69	29620	0.318	BB
15	30.67	5120	0.068	BB

D-7000 HSM: SAMPLES		Series: 0029	Report: original	System: Sys 1	
16	32.08	23660		0.316	BB
17	36.16	353240		4.720	BB
18	37.44	12460		0.166	BB
19	38.88	934600		12.488	BB
20	41.79	84072		1.123	BV
21	46.00	518305		6.925	VV
22	46.65	221091		2.954	VV
23	51.71	341287		4.560	VV
24	52.43	165385		2.213	VV
25	53.41	195974		2.619	VV
26	54.37	282021		3.768	VV
27	55.57	521439		6.967	VV
28	57.23	1011955		13.522	VV
29	57.47	338040		7.436	VV
30	57.97	1252428		17.269	VV
31	59.33	109920		1.369	TBB
		7484012		100.000	

Peak rejection level: 0



3 m H

D-7000 HSM: SAMPLES

Series: 0030

Report: original

System: Sys 1

### D-7000 HPLC System Manager Report

Analyzed: 11/01/00 01:23 PM

Reported: 11/01/00 02:24 PM

Processed: 11/01/00 02:24 PM

Data Path: C:\WIN32APP\HSM\samples\DATA\0030\

Processing Method: tmh

System(acquisition): Sys 1

Series: 0030

Application: SAMPLES

Vial Number: 10

Sample Name: tmh010


Vial Type: UNK

Injection from this vial: 1 of 1

Volume: 20.0 ul

Sample Description:

Chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

Column Type:

Developed by: ratna

Pump A Type: L-7100

Solvent A: air/acet

Pump B Type: L-7100

Solvent B: meth/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc 1	BC
1	1.57	8226	0.157	BV
2	2.21	59211	1.133	VV
3	2.86	51822	0.992	VB
4	3.01	15292	0.254	BV
5	3.33	14108	0.270	VB
6	3.57	81040	1.551	BB
7	5.23	9600	0.184	BB
8	11.97	7520	0.144	BB
9	20.11	9920	0.190	BB
10	23.49	14360	0.275	BB
11	32.85	39720	0.760	BB
12	36.15	63400	1.214	BB
13	37.52	19560	0.374	BB
14	38.93	34280	0.656	BB
15	42.08	67855	1.299	BV

D-7000 HSM: SAMPLES	Series: 0090	Report: original	System: Sys 1
16	43.71	46898	0.898 vv
17	46.13	397776	7.613 vv
18	47.87	86388	1.654 vv
19	49.04	101370	1.944 vv
20	53.75	643226	12.313 vv
21	56.27	877953	16.806 vv
22	56.56	133161	2.932 vv
23	57.33	553252	10.591 vv
24	57.80	319370	6.114 vv
25	58.08	1441325	27.591 vv
26	59.44	109040	2.087 TBB
		5223893	100.000

Peak rejection level: 0



D-7000 HSM: SAMPLES

Series: 0050

Report: original

System: Sys 1

Channel 1 Noise: Not Measured  
Channel 1 Drift: Not Measured

Configuration parameters:

Interface module: D-7000  
Channel 1 Detector: L-7450  
Column Oven: None  
Pump A: L-7100  
Pump C: None  
Column Name:

Gradient Mode: High  
Channel 2 Detector: None  
Autosampler: Manual  
Pump B: L-7100  
External Instrument Software: None

Method Information:

Method Name: tmin  
Description:  
analisaDAG

Developed by: ratna

Pump Setup:

All Pumps Pressure Limit: 0 to 420 kgf/cm<sup>2</sup>

Pump A (L-7100):

Solvent A: air/acet

Pump B (L-7100):

Solvent B: meoh/acet

Pump Table

Time (min)	%air/a	%meoh/	Total Flow (ml/min)	Event 1	Event 2	Event 3	Event 4
0.0	100.0	0.0	1.000				
5.0	100.0	0.0	1.000				
55.0	0.0	100.0	1.000				
60.0	100.0	0.0	1.000				

Channel 1 Detector (L-7450):

Spectral Bandwidth: 4 nm  
Absorbance Mode: Normal (2.0AU)  
Wavelength Range: 190 to 390 nm  
Start Time: 0.00 min

Spectral Interval: 1500 msec  
Auto Zero before Injection: YES  
Monitoring Wavelength: 260 nm  
Stop Time: 60.00 min

Method DP for channel 1

Calculation Method:

Calculation Method: AREA3  
Peak Identification Window: % Time  
UNK peaks identification rule: Closest peak  
Update RT in component Table: NO  
Do library search: NO

Peak Quantitation: Area  
Concentration data from method.

Component Table

RT (min)	Window (%)	Func1	Func2	Func3	El-Conc	Tolerance (%)
5.00	10.00					

Integration Table

time (min)	function	value/status
0.00	NOISE	10
0.00	SMOOTHING	OFF
0.00	SENSITIVITY	20
0.00	N-METHOD	0

## DAD Processing setup:

Peak spectrum integration enabled: NO  
 Chromatogram to create: Fixed at 260 nm

Peak purity check enabled: NO  
 Delete DAD data after reporting: NO

## DAD Display Format:

Time range: 0.00 to 60.00 min  
 Offset: 0.0  
 Auto Mark Peak WL: NO  
 3-D resolution: Medium  
 3-D rotation: 30  
 Display spectra only: NO

Absorbance scale (Au): Auto  
 Wavelength range: 190 to 390 nm  
 Spectrum display: Absorbance  
 Auto BG Subtraction: NO  
 3-D tilt: 30  
 3-D mirror: NO  
 Report spectra: Peak top and sides.

Perform system suitability test : NO  
 Perform module performance test : NO  
 Perform data diagnosis : NO

## Chromatogram Display Format:

Autoscale Time Range: 0.00 to 60.00 min  
 Use alternate scale: NO  
 Scale to Full Chrom Time Range: YES  
 Baseline overlay: YES  
 Marker-in signals: NO  
 Show integration time table: NO  
 Show gradient curves: NO  
 Report channel 1 labels in the chromatogram overlay graph.  
 Picture in picture: none

Autoscale: YES  
 Auto zero: NO  
 Peak rejection level: 0 uV \* sec  
 Peak start-end markers: YES  
 Peak labels: time, none  
 Show sampling period time table: NO

## Report Format:

Use primary layout: YES  
 Print primary layout report: YES  
 Acquisition DDE: NO  
 Reprocess DDE: NO  
 Concentration 1 unit: Other  
 Concentration 1 scale factor: 1.0000  
 Concentration 1 divide by sample amount: NO  
 Concentration 2 unit: other  
 Concentration 2 scale factor: 1.0000  
 Concentration 2 use component multiplier: NO  
 Injection report column 1 header: PK-NUM  
 Injection report column 2 header: RT  
 Injection report column 3 header: AREA  
 Injection report column 4 header: CONC1  
 Injection report column 5 header: BC

Reported peaks: All Peaks  
 Use secondary layout: NO  
 Print secondary layout report: NO  
 Acquisition macro name:  
 Reprocess macro name:  
 Concentration 1 name:  
 Concentration 2 name:

7/11/17

D-7000 HSM: SAMPLES

Series: 0031

Report: original

System: Sys 1

### D-7000 HPLC System Manager Report

Analyzed: 11/01/00 02:29 PM

Reported: 11/01/00 03:29 PM

Processed: 11/01/00 03:29 PM

Data Path: C:\WIN92APP\HSM\samples\DATA\0031\

Processing Method: tmh

System(acquisition): sys 1

Series: 0031

Application: SAMPLES

Vial Number: 1

Sample Name: tmh001


Vial Type: UNK

Injection from this vial: 1 of 1

Volume: 20.0 ul

Sample Description:

Chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

Column type:

developed by: ratna

Pump A Type: L-7100

Solvent A: air/acet

Pump B Type: L-7100

Solvent A: mech/acet

Peak Quantitation: AREA

Calculation Method: AREA%

NO.	RT	Area	conc 1	BC
1	2.91	95440	1.796	BB
2	3.04	20840	0.392	BV
3	3.33	15560	0.293	VB
4	3.57	44880	0.845	BB
5	4.12	20080	0.378	BB
6	12.40	18050	0.340	BB
7	13.17	22360	0.421	BB
8	16.91	30029	0.565	BB
9	33.07	38660	0.727	BB
10	36.37 ✓	64100	1.206	BB
11	37.65 ✓	13350	0.251	BB
12	39.09	32285	0.608	BB
13	43.84	109685	2.064	BV
14	46.40	397260	7.475	VV
15	46.03	104266	1.962	VV

D-7000 HSM: SAMPLES	Series: 0051	Report: original	System: Sys 1
16	49.07	96934	1.624 VV
17	49.92	91340	1.719 VV
18	55.67	602191	11.531 VV
19	54.80	309820	5.830 VV
20	56.35	592780	11.154 VV
21	57.41	723642	13.617 VV
22	57.65	523657	9.654 VV
23	58.16	1254260	23.602 VV
24	59.49	92840	1.747 TBB
		5314313	100.000

Peak rejection level: 0





D-7000 HSM: SAMPLES

Series: 0031

Report: original

System: Sys 1

Channel 1 Noise: Not Measured  
Channel 1 drift: Not Measured

Configuration parameters:

interface module: D-7000  
channel 1 detector: L-7450  
Column Oven: None  
Pump A: L-7100  
Pump C: None  
column name:

gradient mode: high  
channel 2 detector: none  
Autosampler: Manual  
Pump B: L-7100  
External Instrument Software: None

Method Information:

Method Name: tmh  
Description:  
analisaD&S

Developed by: ratna

Pump Setup:

All Pumps Pressure Limit: 0 to 420 kgf/cm<sup>2</sup>

Pump A (L-7100):

solvent A: air/acet

Pump B (L-7100):

Solvent B: meoh/acet

Pump table

time (min)	%air/a	%meoh/	Total Flow (ml/min)	Event 1	Event 2	Event 3	Event 4
0.0	100.0	0.0	1.000				
5.0	100.0	0.0	1.000				
55.0	0.0	100.0	1.000				
60.0	100.0	0.0	1.000				

Channel 1 Detector (L-7450):

Spectral Bandwidth: 4 nm  
Absorbance Mode: Normal (2.0AU)  
wavelength range: 190 to 890 nm  
start time: 0.00 min

Spectral Interval: 1600 msec  
Auto Zero before Injection: YES  
monitoring wavelength: 260 nm  
stop time: 60.00 min

Method DP for channel 1

calculation method:

Calculation method: AREAS

Peak identification window: % Time

UNK peaks identification rule: Closest peak

Update RT in component Table: NO

no library search: NO

Peak quantitation: Area  
concentration data from method.

Do blank subtraction: NO

Component Table

RT (min)	Window (%)	Func1	Func2	Func3	E-Conc	Tolerance (%)
5.00	10.00					

NO HSM: SAMPLES

Series: 0031

Report: original

System: Sys 1

me	function	value/status
00	NOISE	10
00	SMOOTHING	OFF
00	SENSITIVITY	20
00	N-METHOD	0

processing setup:  
peak spectrum integration enabled: NO  
chromatogram to create: Fixed at 250 nm

peak purity check enabled: NO  
delete DAD data after reporting: NO

Display Format:  
time range: 0.00 to 60.00 min  
offset: 0.0  
Auto Mark Peak WL: NO  
X-D resolution: Medium  
X-D rotation: 30  
display spectra only: NO

Absorbance scale (Au): Auto  
wavelength range: 190 to 390 nm  
spectrum display: Absorbance  
Auto BG Subtraction: NO  
3-D tilt: 30  
3-D mirror: NO  
report spectra: peak top and sides.

form system suitability test : NO  
form module performance test : NO  
form data diagnosis : NO

Chromatogram Display Format:  
autoscale Time Range: 0.00 to 60.00 min  
use alternate scale: NO  
scale to Full Chrom Time Range: YES  
baseline overlay: YES  
marker-in signals: NO  
show integration time table: NO  
show gradient curves: NO  
report channel 1 labels in the chromatogram overlay graph.  
picture in picture: none

Autoscale: YES  
Auto zero: NO  
Peak rejection level: 0 uV \* sec  
peak start-end markers: YES  
peak labels: time, none  
Show sampling period time table: NO

Report Format:  
use primary layout: YES  
print primary layout report: YES  
acquisition DDE: NO  
reprocess DDE: NO  
concentration 1 unit: Other  
concentration 1 scale factor: 1.0000  
concentration 1 divide by sample amount: NO  
concentration 2 unit: other  
concentration 2 scale factor: 1.0000  
concentration 2 use component multiplier: NO  
injection report column 1 header: PK-NUM  
injection report column 2 header: RT  
injection report column 3 header: AREA  
injection report column 4 header: CONC1  
injection report column 5 header: BC

Reported peaks: All Peaks  
use secondary layout: NO  
Print secondary layout report: NO  
Acquisition macro name:  
Reprocess macro name:  
Concentration 1 name:  
concentration 2 name:

### D-7000 HPLC System Manager Report

lyzed: 11/02/00 08:20 AM

Reported: 11/02/00 10:45 AM

Processed: 11/02/00 09:20 AM

Path: C:\WIN52APP\HSM\samples\DATA\0032\

Processing Method: tmh

tem(acquisition): Sys 1

Series: 0032

Location: SAMPLES

Vial Number: 10

File Name: tmh010


Vial Type: UNK

Injection from this vial: 1 of 1

Volume: 20.0 ul

File Description:

Chrom type: fixed WL chromatogram, 260 nm



Acquisition Method: tmh

Column type:

developed by: ratna

Column A Type: L-7100

Solvent A: air/acet

Column B Type: L-7100

Solvent B: meth/acet

Peak Quantitation: AREA

Integration Method: AREA%

RT	Area	conc 1	BC
2.91	103320	1.132	BB
3.07	50761	0.556	BV
3.55	629339	6.858	VV
9.25	20480	0.224	TBB
11.79	39758	0.433	VB
14.99	4500	0.049	BB
16.43	9880	0.108	BB
18.53	4760	0.052	BB
31.15	22680	0.249	BB
33.95	96400	1.057	BB
36.48	418803	4.589	BV
37.79	6120	0.067	TBB
39.25	1022679	11.209	VB
42.11	123301	1.351	BV
46.53	819168	8.979	VV

D-7000 HSM: SAMPLES		Series: 0032	Report: original	System: Sys 1	
16	48.05	109015		1.195	VV
17	49.17	245768		2.694	VV
18	50.05	159980		1.754	VV
19	50.89	88940		0.758	VV
20	51.87	351531		3.853	VV
21	52.56	283441		3.107	VV
22	53.66	215558		2.562	VV
23	56.00	1287085		14.108	VV
24	57.44	882637		9.874	VV
25	57.68	406061		4.451	VV
26	58.13	1723386		18.890	VV
		9123331		100.000	

Peak rejection level: 0

